


# PRODUCT DATA SHEET

Revision 1	<h2 style="margin: 0;">Cell Culture Tube</h2>	
Validfor Item-No.:	1120000/1121000/1125000	

1.	Description / Specification	
1.1	Description	Cell Culture Tube, filter cap  1120000 Cell Culture Tube, 15 ml, sterile 1121000 Cell Culture Tube, 50 ml, sterile 1125000 Cap of Cell Culture Tube, 50 ml, sterile
1.2	Dimensions	Flask: see Customer Drawing Pore size of filter membrane: 0.22 µm
1.3	Volume	Max volume: 15ml (1120000) /50ml (1121000) Work volume:3-10ml (1120000) /10-30ml (1121000)
1.4	Material / Resin	Tube: PS (Polystyrene) Cap: HDPE (High Density Polyethylene) Filter: PVDF (Polyvinylidene fluoride) The materials for manufacturing are <i>USP Class VI</i>
1.5	Colour	Tube: clear Cap: blue
1.6	Sterilization	SAL 10 <sup>-6</sup>
1.7	Quality Control	Raw Material-Control: physical testing Product-Control: testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for cell culture to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only Expiry date and Lot-No. printed on the outer box

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens. Contents non-cytotoxic
2.2	Temperature range	For application: -80°C to +121 °C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Shelf life	3 years
2.6	Other Information	-

3.	Packaging	1120000	1121000	1125000
3.1	Pieces / Bag	25	25	25
3.2	Pieces / Box	20	8	40
3.3	Lot-No.	DDD YY XXX (day, year, product No.)		
3.4	Other Information	Certificate of Quality to download		

4.	Other Information
4.1	Laboratory use only

Prior Issue	Drawn	Approved	Released	<b>CONFIDENTIAL:</b> Information contained in this document or drawing is confidential and proprietary to Zhejiang Saining Biotechnology Co., Ltd. This document may not be reproduced for any reason without written permission from Zhejiang Saining Biotechnology Co., Ltd. All rights of design, invention, and copyright are reserved.
Revision 1	Date 1 June 2024	Date 1 June 2024	Date 1 June 2024	
Date 01.06.2024	Name Chen Anjing	Name Zhang Heng	Name Wang Kedong	

**DISCLAIMER:** The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.